

**Measurement of UV radiation from exhaust and suction SRSmedilux: PMX2Axx, AW48**



The subject of the measurement is harmful UV-C radiation invisible to the human eye (blue radiation visible to the eye is not harmful to humans and living organisms).


The measurement was performed by using of the calibrated spectrometer **ASEQ INSTRUMENTS LR1-B V1.0** at first in the opening of the UVC chamber, then in the exhaust / suction of PMX2Axx, AW48.

Opening to the UVC chamber after screwing out one screw.

**Chamber UVC measurement:**



Probe aimed at the opening of the UVC chamber.




Measured peak of UVC radiation.

**Exhaust / suction UVC measurement:**



Probe aimed at the exhaust / suction.



UVC radiation was not detected.

Bratislava, 1.10.2020

SRS Group s.r.o.  
Rybnicna 36/D, 831 07 Bratislava, Slovakia  
Phone: +421948590709, email: sales@srs-group.com  
IČO: 44520433, IČ-DPH/VAT No.: SK2022726629

